

Specification: AX 500

The logo features the word "COMEN" in a large, bold, 3D-style font. The letters are white with a blue gradient and a shadow effect. Above the letters, there is a decorative arrangement of smaller "COMEN" text elements in various sizes and orientations, creating a cloud-like or particle effect. The background is a solid blue color with a subtle gradient and a white curved line at the bottom, suggesting a globe or a horizon.

COMEN Share with the world

Anesthesia Machine

AX 500



=====Technical Specification=====

Physical Characteristics

Size	678 mm × 580 mm × 1370 mm
Weight	90kg
Maximum Bearing	
Weight	160kg
Screen Size:	10.4" TFT touch screen
Resolution	800 × 600
Handrail Length	412mm
Caster wheel	4 wheels 5" brakes;

Operation Environment

Working Temp	10~40°C
Humidity	≤93%
Power Supply	100-240V~, 50/60Hz±1Hz
Battery Type	Rechargeable Lithium-ion battery
Battery Capacity	4400mAh, 11.1VDC
Battery Recharging	
Time	4 hours for charging
Battery backup Trace	2 hours for continuous working
Waveforms:	Pressure-time; Flow rate-time; Capacity-time; ET EtCO2 concentration;
Optional:	Pressure-volume Loops; Flow-volume Loops; Pressure- flow Loops

Top Plate

Maximum supporting capacity	50kg
Operational dimensions	535mm×235mm
Dimensions with Additional Accessory	508mm×313mm×380mm

Workbench

Maximum supporting capacity	20kg
-----------------------------	------

Operational dimensions	465mm×275mm
Dimensions with Additional Accessory	472mm×248mm×380mm

Interface:

- USB port
- RJ45
- 3 auxiliary power output
- AC power interface
- Equal-potential grounding terminal
- DB9 interface

Features

Drawers:	Size:416mm×395mm×170mm
	Bearing Weight: 1Kg
	Length: 320mm; Height: 240mm

Gas-bag Sway

Brace

Anesthesia process	Open, semi closed, closed circuit
Patients	Adult, pediatric
Working Mode	Manual, Mechanical, Standby
Compliance	Compliance Correction
Configuration	Possibility of configuration observation
Bypass	Available
Optional	Heating; Oxygen sensor; ACGO; AGSS; MASIMO EtCO2 (sidestream); MASIMO EtCO2 (mainstream); MASIMO AG (sidestream); Respironics EtCO2 (mainstream); BIS; CPB

=====Ventilator Specification=====

Ventilation Modes

VCV/VC	Volume-Controlled Ventilation with tidal volume compensation
PCV/VPC	Pressure Control Ventilation
Others	Manual and automatic ventilation
Optional	SIMV-VC, PSV/ CPAP, SIMV-PC, PRVC, SIMV-PRVC, PSVPro
Ventilation principle	Chronometric, volumetric and barometric
Ventilation	Electronically controlled& pneumatically driven
Driven gas	O ₂ (air: optional)
Breathing circuit volume	1000 ml + bag

Ventilator Setting ranges

Monitoring parameter	Tidal volume, Inspiratory, expiratory flow, minute volume, frequency, pressure (P _{mean} , P _{plat} , P _{peak} , PEEP), Oxygen, CO ₂ , N ₂ O and halogenated expiratory concentration, Pressure, oxygen, CO ₂ , N ₂ O and Halogen numerical values, compliance and patient resistance
Tidal volume range	15 ~1500 mL
MV (Per-minute ventilation amount)	0~100 L/min
Pressure range (limit)	10~100 cmH ₂ O
Pressure range (support)	3~60cmH ₂ O
Respiratory rate	4~100bpm
Inspiratory /Expiratory ratio (I: E) range	4:1~1:10
Apnea I: E	4:1~1:8
Apnea time	10~30s
Apnea pressure	3~60cmH ₂ O
Freq. Min. (Min. frequency for apnea-ventilation)	2-60 bpm
Inspiratory pause	OFF, 5~60% of inspiratory time
Inspiratory time	0.2~5s
Inspiratory pressure	5~70cmH ₂ O
PEEP	OFF, 3~30cmH ₂ O
Trigger pressure	-20~-1cmH ₂ O
Trigger window	5~90%

Trigger flow	0.2~15 L/ min
Flush oxygen	25~75 L/ min
Inspiratory stop level	5~80%
Pressure slope	0~2.0s

Ventilator Monitoring Ranges

TV (Inspiratory tidal volume)	0~3000 mL
TV (Expiratory tidal volume)	0~3000 mL
MV (Per-minute ventilation amount)	0~100 L/min
FiO ₂ (Oxygen concentration)	18~100%
Airway pressure	-20~120cmH ₂ O
PEEP	0~70cmH ₂ O
P _{peak} (Airway pressure)	-20~120 cmH ₂ O
P _{mean} (Mean pressure)	-20~120cmH ₂ O
P _{plat} (Platform pressure)	0~120cmH ₂ O
I: E (Inspiratory-expiratory ratio)	4:1~1:12
Freq (Respiratory rate)	0~120 bpm
Compl (Compliance)	0~300 mL/cmH ₂ O
Resistance	0~600 cmH ₂ O/(s/L)
EtCO₂	
MASIMO EtCO ₂ (sidestream);	0~190mmHg, 0~25% (at 760mmHg) Accuracy: ± (0.3%+4% of reading).
MASIMO EtCO ₂ (mainstream)	0~190mmHg, 0~25% (at 760mmHg) Accuracy: ± (0.3%+4% of reading).
Respironics EtCO ₂ (mainstream)	0~150mmHg, 0~19.7% (at 760mmHg) Accuracy: 0~5.3%: ±0.3%; 5.4~9.2%: ±5% of reading; 9.3~13.2%: ±8% of reading; 13.3~19.7%: ±10% of reading;
AG	
MASIMO AG	SEV: 0~25% DES: 0~25% HAL/ ISO/ ENF: 0~25% N ₂ O: 0~100% O ₂ : 0~100% CO ₂ : 0~25% (0~190mmHg)

Accuracy:
 SEV: 0~1%: $\pm 0.15\%$; 1~5%: $\pm 0.2\%$;
 5~8%: $\pm 0.4\%$;
 DES: 0~1%: $\pm 0.15\%$; 1~5%: $\pm 0.2\%$;
 5~10%: $\pm 0.4\%$; 10~15%: $\pm 0.6\%$;
 15~18%: $\pm 1\%$;
 ISO, ENF, HAL: 0~1%: $\pm 0.15\%$;
 1~5vol %: $\pm 0.2\%$;
 N2O: $\pm (2\% + 2\%$ of the reading)
 O2: 0~25%: $\pm 1\%$; 25~80%: $\pm 2\%$;
 80~100%: $\pm 3\%$;
 CO2: 0~15%: $\pm (0.2\% + 2\%$ of the reading); 15~25%: unspecified

Anesthesia depth

BIS 0.0~100.0
 SQI 0.0~100.0%
 EMG 0~100dB
 ESR 0.0~100.0%

Ventilator Performance

Pressure range at inlet 0.28~0.6 MPa
 Peak gas flow >100 L/min
 Flow valve range 1~100 L/min
 Flow compensation range 200 mL/min to 18 L/min
 Inspiratory flow Maximum inspiratory flow shall not be smaller than 120L/min when gas supply pressure is 280KPa.
 Range of flow valve 3~100 L/min
 Pressure limitation Controlled by the electronic relief valve fitted inside the ventilator;
 Controlling means for ventilator Controlled by the mechanical relief valve fitted inside the ventilator.

=====Ventilator accuracy=====

Control accuracy

TV 15~60 ml: $\pm 10\text{ml}$;
 60~210 ml: $\pm 15\text{ml}$;
 210~1500 ml: $\pm 7\%$ of set value.
 PCV Inspiratory pressure: $\pm 2.5\text{cmH}_2\text{O}$ or $\pm 7\%$ of set value, whichever the greater.
 Limiting pressure: $\pm 2.5\text{cmH}_2\text{O}$ or $\pm 7\%$ of set value, whichever the greater.

PEEP: OFF: undefined; 3~30cmH₂O: $\pm 2.0\text{cmH}_2\text{O}$, or $\pm 8\%$ of set value, whichever is the greater.
 Supporting pressure: $\pm 2.5\text{cmH}_2\text{O}$ or $\pm 7\%$ of set value, whichever the greater.
 Apnea pressure: $\pm 2.5\text{cmH}_2\text{O}$ or $\pm 7\%$ of set value, whichever the greater.
 Trigger pressure: $\pm 2.0\text{cmH}_2\text{O}$.
 Freq ± 1 bpm or $\pm 5\%$ of set value, whichever is the greater.
 I: E 2: 1~1: 4: $\pm 10\%$ of reading value;
 Other ranges: $\pm 25\%$ of reading value.
 Apnea I: E 2: 1~1: 4: $\pm 10\%$ of set value;
 Other ranges: $\pm 25\%$ of set value.
 Tpause In the range of 20%~60%: $\pm 15\%$ of set value;
 Other ranges: undefined.
 Inspiratory time $\pm 0.2\text{s}$
 Inspiratory pause 20%~60%: $\pm 15\%$ of set value;
 Other ranges: undefined.
 Trigger window $\pm 10\%$
 Trigger flow rate ± 1 L/ min
 Inspiratory stop level $\pm 10\%$
 O₂/ N₂O/ Air flow control 10~100% of the full scale: $\pm 10\%$ of the reading value. Other ranges: undefined.
 Total flow control Air balance gas: $\leq \pm 3\%$
 N₂O balance gas: $\leq \pm 3\%$
 Backup flow control Pure Oxygen flow rate is 0~10 L/min: $\leq \pm 3\%$; Others: undefined.
 Auxiliary flow control 10~100% of the full scale: $\pm 10\%$ of the reading value. Other ranges: undefined.

Monitoring accuracy

TV (expiratory) 0~60ml: ± 10 ml; 60ml ~ 3000ml: $\pm 20\text{ml}$ or $\pm 7\%$ of reading value, whichever is greater; Others: undefined.
 TV (Inspiratory) 60ml ~ 3000ml: $\pm 20\text{ml}$ or $\pm 7\%$ of reading value, whichever is greater; Others: undefined.
 Paw -20 cmH₂O~120 cmH₂O: ± 2.0 cmH₂O or $\pm 4\%$ of set value, whichever is greater; Others: undefined.

PEEP	0 cmH ₂ O~70 cmH ₂ O: ±2.0 cmH ₂ O or ± 4% of set value, whichever is greater; Others: undefined.
Pmean	-20 cmH ₂ O~120 cmH ₂ O: ±2.0 cmH ₂ O or ± 4% of setting value, whichever is greater; Others: undefined.
Pplat	0 cmH ₂ O~120 cmH ₂ O: ±2.0 cmH ₂ O or ± 4% of set value, whichever is greater; Others: undefined.
Freq	±1 bpm or ±5% of set value, whichever is the greater.
I: E	2: 1~1: 4: ±10% of reading value; 4: 1~2: 1 and 1: 4~1: 12: ±25% of setting value; Others: undefined.
MV	0 L/min~30 L/min: ±1 L/min or ±15% of set value, whichever is greater; Others: undefined.
Compliance	0 ml/cmH ₂ O~250 ml/cmH ₂ O: ±0.5 ml/cmH ₂ O or ± 15% of reading value, whichever is greater; Other ranges: undefined.
Resistance	0 cmH ₂ O/(L/s)~20 cmH ₂ O/(L/s): ±10 cmH ₂ O/(L/s); 20 cmH ₂ O/(L/s)~500 cmH ₂ O/(L/s): ±50% of reading value; Other ranges: undefined.
Oxygen sensor	±3%
O ₂ / N ₂ O/ Air flow control	10~100% of the full scale: ±10% of the reading value. Other ranges: undefined.
Total flow control	Air balance gas: ≤±3% N ₂ O balance gas: ≤±3%
Backup flow control	Pure Oxygen flow rate is 0~10 L/min: ≤±3%; Others: undefined.
Auxiliary flow control	10~100% of the full scale: ±10% of the reading value. Other ranges: undefined.
Alarm Settings	
Tidal volume (expiratory)	High: 5~1600 ml Low: 0~1595 ml
MV	High: 2~100L/min Low: 0~98L/min
Inspired oxygen	High: 20~105% Low: 18~103%

Ppeak	High: 2~100cmH ₂ O Low: 0~98cmH ₂ O
Apnea alarm	Two (2) triggering conditions are satisfied simultaneously: 1. Airway pressure is continuously lower than (PEEP +3) cmH ₂ O for more than 30 seconds. 2. Expiratory tidal volume is continuously lower than 10ml for more than 30 seconds. Increase the set values of tidal volume and respiratory frequency, or set it to Manual/spontaneous mode. Audible and visual alarm;
Alarm	Audible and visual alarm;
Alarm access	Easy access by shortcut
Flow meters	
Type	Mechanical flow meter
Gas Supply	
Pipeline gasses	O ₂ , Air
Optional	O ₂ , Air, N ₂ O; O ₂ , N ₂ O
Backup	
gas-cylinder gasses	O ₂ , N ₂ O, Air
Pipeline gas connection	NIST
Backup cylinder connection	YOKE-CGA
Pressure range at inlet	280~600 kPa
Filter	60-80um
Features	Switch easily to the other gas without interrupting the ventilation
Auxiliary gas supply	O ₂ (optional)
=====Breathing Circuit Specification=====	
System Pressure Gauge	
Range	-20~100 cmH ₂ O
Accuracy	± (4% of full scales reading + 4% of reading)
Adjustable Pressure Limiting (APL)valve	
Range	1~75 cmH ₂ O
Tactile knob indication at	>30 cmH ₂ O
Accuracy:	±1.0 cmH ₂ O
Minimum opening	

pressure 0.3 cmH₂O (dry), 0.5 cmH₂O (humid)

Breathing Circuit Parameters

Compliance ≤4mL/100Pa
Automatically compensates for compression loss with in the breathing circuit in mechanical mode

Volume of CO₂ canister 2000ml

Feature Heated at 134 degree, removable, easy to dismantle and sterilize

=====**Gas Monitoring**=====

Carbon Dioxide (CO₂) Modules

Type Mainstream ETCO₂, Sidestream ETCO₂
Method Infrared absorption
Display Numeric and curve displayed in screen
Alarm delay 1~10s (step size: 1s)
Sweep 6.25 mm/s,12.5 mm/s

Anesthetic Agent (AG) Module

Maximum sound pressure for low alarm 79dB
Measurement type Side stream
Module type Phasin ISA AG module
Accuracy ±10ml/min or ±10%, whichever is greater
Monitored parameters CO₂, N₂O, AA, MAC, Paramagnetic O₂ and BIS

Active AGSS

Feature High flow, low vacuum
Size 535mm×120mm×155mm
Weight 2.2kg
Applies ISO 80601-2-13 and YY 0635-2

Pressure relief device Atmospheric pressure compensation port

Connector ISO9170-2 or BS6834 standard connector

Flow of suction 50-80L/min

Resistance 0.75KPa ,75L/min

Filter Stainless steel mesh, with pore size of 60~100μm

ACGO

Connector Taper coaxial fitting of 22mm (outside) and 15 (inside)

Back pressure generated at the rear end of anesthesia vaporizer and the front-end of ACGO during quick oxygen charging ≤2kPa

Flush O₂

100% fast oxygen

Vaporizer

Brand Drager and Penlon available
Locking Vaporizer with interlocking system (Optional: Two vaporizers)
Automatic recognition Anesthesia machine able to automatic recognize halogenated gases

=====**Power(No isolation transformer)**=====

External AC power supply

Input voltage 100~240 V~/ 100~120V~

Input current 3.5~8.5 A/8.5 A

Input frequency 50/60 Hz

Leakage current < 500μA

Auxiliary output supply

Output voltage 100~240 V~/ 100~120V~

Output frequency 50/60 Hz

***Notice: Specifications subject to changes without prior notice. All rights reserved by Comen**